



| 波長 (nm) | RawData ... |
|---------|-------------|
| 300.00 | 0.0168 |
| 310.00 | 0.0214 |
| 320.00 | 0.0188 |
| 330.00 | 0.0152 |
| 340.00 | 0.0069 |
| 350.00 | 0.0083 |
| 360.00 | 1.9957 |
| 370.00 | 22.7949 |
| 380.00 | 52.6939 |
| 390.00 | 71.7702 |
| 400.00 | 82.2337 |
| 410.00 | 87.8835 |
| 420.00 | 90.9590 |
| 430.00 | 92.7224 |
| 440.00 | 93.6652 |
| 450.00 | 94.3645 |
| 460.00 | 94.8653 |
| 470.00 | 95.0862 |
| 480.00 | 95.3804 |
| 490.00 | 95.5656 |
| 500.00 | 95.7434 |
| 510.00 | 95.8318 |
| 520.00 | 95.9823 |
| 530.00 | 96.0544 |
| 540.00 | 96.1733 |
| 550.00 | 96.3849 |
| 560.00 | 96.4830 |
| 570.00 | 96.5559 |
| 580.00 | 96.6273 |
| 590.00 | 96.6589 |
| 600.00 | 96.6894 |
| 610.00 | 96.6484 |
| 620.00 | 96.4537 |
| 630.00 | 96.2914 |
| 640.00 | 95.9918 |
| 650.00 | 95.5958 |
| 660.00 | 95.1009 |
| 670.00 | 94.5193 |
| 680.00 | 93.7299 |
| 690.00 | 92.9827 |
| 700.00 | 92.0032 |
| 710.00 | 91.0268 |
| 720.00 | 89.9339 |
| 730.00 | 88.8346 |
| 740.00 | 87.3822 |
| 750.00 | 86.2177 |
| 760.00 | 84.8538 |
| 770.00 | 83.1131 |
| 780.00 | 81.6242 |
| 790.00 | 80.2670 |
| 800.00 | 78.7694 |
| 810.00 | 77.3442 |
| 820.00 | 75.8619 |
| 830.00 | 74.4848 |
| 840.00 | 73.1187 |
| 850.00 | 71.6780 |
| 860.00 | 70.3585 |
| 870.00 | 69.0319 |
| 880.00 | 67.7074 |
| 890.00 | 66.4526 |
| 900.00 | 65.2518 |
| 910.00 | 64.0925 |
| 920.00 | 62.9460 |
| 930.00 | 61.8496 |
| 940.00 | 60.8431 |
| 950.00 | 59.8036 |
| 960.00 | 58.8190 |
| 970.00 | 57.8705 |
| 980.00 | 56.9927 |
| 990.00 | 56.1811 |
| 1000.00 | 55.3932 |
| 1010.00 | 54.6142 |
| 1020.00 | 53.8699 |
| 1030.00 | 53.2213 |
| 1040.00 | 52.5886 |
| 1050.00 | 51.9299 |
| 1060.00 | 51.3357 |
| 1070.00 | 50.7709 |
| 1080.00 | 50.2270 |
| 1090.00 | 49.7226 |
| 1100.00 | 49.2329 |
| 1110.00 | 48.7578 |
| 1120.00 | 48.3457 |
| 1130.00 | 47.9270 |
| 1140.00 | 47.5407 |
| 1150.00 | 47.1743 |
| 1160.00 | 46.8204 |
| 1170.00 | 46.4848 |
| 1180.00 | 46.1642 |
| 1190.00 | 45.8750 |
| 1200.00 | 45.6122 |

[測定プロパティ]
 波長範囲(nm): 300.00 ~ 1300.00
 スキャン スピード: 中速
 サンプルングピッチ: 0.5
 オート サンプルングピッチ: 無効
 測定モード: オート

[装置プロパティ]
 装置の種類: UV-3600Plus シリーズ
 測光値: 透過率
 スリット幅: (20)
 時定数: 0.1 秒
 光源: 自動
 光源切替波長: 282.00 nm
 検出器ユニット: 外部(3検出器)
 検出器切替波長: 780.00 nm 1800.00 nm